



ADMINISTRATIVE RULES OF MONTANA



37.14.1021 SPECIAL REQUIREMENTS FOR CHEST PHOTOFLUOROGRAPHIC INSTALLATIONS

- (1) The equipment used in chest photofluorographic installations shall be as follows:
 - (a) All requirements of ARM 37.14.1012 (1) shall be satisfied.
 - (b) A collimator shall restrict the useful beam to the area of the photofluorographic screen.
- (2) All requirements of ARM 37.14.1003 (5) and 37.14.1012 (2) concerning structural shielding shall be satisfied.
- (3) The operating procedures applicable to chest photofluorographic installations shall be as follows:
 - (a) All requirements of ARM 37.14.1012 (3) shall be satisfied.
 - (b) All individuals except the patient being examined shall be in shielded positions during exposures.
 - (c) Personnel monitoring shall be required for all individuals associated with the operation of the equipment.

Authorizing statute(s): Sec. 75-3-201 and 75-3-202, MCA

Implementing statute(s): Sec. 75-3-201, 75-3-202 and 75-3-204, MCA

History: NEW, 1980 MAR p. 1069, Eff. 3/28/80; TRANS, from DHES, 1996 MAR p. 433; TRANS, from DEQ, 2000 MAR p. 189.